



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG No.: O2213
DataFile: FF012422.D Analyst Name: YP\AJ Analyst Date: 04-08-2023

Conc. (PPM)	Area Count	RF	Average RF	%D
500	57842499	115685	126682	8.681