



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Instrument ID: ECD_R Calibration Date(s): 03/13/2023 03/13/2023
Calibration Times: 10:20 10:59

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PR060260.D</u>		CF 0.25 = <u>PR060261.D</u>			
CF 0.1 = <u>PR060262.D</u>		CF 0.05 = <u>PR060263.D</u>		CF 0.025 = <u>PR060264.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	97465700000	110484000000	102833000000	107659000000	101983000000	104085000000	5
EDB	56436600000	64731400000	62341900000	65249000000	64325200000	62616800000	6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Instrument ID: ECD_R Calibration Date(s): 03/13/2023 03/13/2023

Calibration Times: 10:20 10:59

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PR060260.D</u>		CF 0.25 = <u>PR060261.D</u>			
CF 0.1 = <u>PR060262.D</u>		CF 0.05 = <u>PR060263.D</u>		CF 0.025 = <u>PR060264.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	87043000000	97520400000	96300100000	99871900000	100665000000	96280000000	6
EDB	45912900000	52001300000	52732000000	55351500000	53058400000	51811200000	7