



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/06/2023 Initial Calibration Date(s): 03/13/2023 03/13/2023

Continuing Calib Time: 11:38 Initial Calibration Time(s): 10:20 10:59

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DBCP	6.65	6.65	6.55	6.75	0.01
EDB	2.88	2.88	2.78	2.98	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/06/2023 Initial Calibration Date(s): 03/13/2023 03/13/2023

Continuing Calib Time: 11:38 Initial Calibration Time(s): 10:20 10:59

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DBCP	5.41	5.41	5.31	5.51	0.00
EDB	2.16	2.15	2.05	2.25	-0.01