



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/13/2023 03/13/2023

Client Sample No.: CCAL01 Date Analyzed: 04/06/2023

Lab Sample No.: M8011-504 CCC 0. Data File : PR060684.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	6.645	6.545	6.745	0.090	0.100	0.0
EDB	2.879	2.779	2.979	0.090	0.100	0.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/13/2023 03/13/2023

Client Sample No.: CCAL01 Date Analyzed: 04/06/2023

Lab Sample No.: M8011-504 CCC 0. Data File : PR060684.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.407	5.310	5.510	0.090	0.100	0.0
EDB	2.155	2.054	2.254	0.100	0.100	0.0