



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR031323

Review By	Ankita	Review On	3/14/2023 9:13:04 AM
Supervise By	Sohil	Supervise On	3/14/2023 9:14:23 AM
SubDirectory	PR031323	HP Acquire Method	HP Processing Method PR031323
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP21791,PP21792,PP21793,PP21794,PP21795		
CCC	PP21798		
Internal Standard/PEM			
ICV/I.BLK	PP21796		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	Data File Name	Date-Time	Operator	Status
1	HEXANE	PR060259.D	13 Mar 2023 10:11	AJMA	Ok
2	M8011.504.1 0.5 PPB ICC	PR060260.D	13 Mar 2023 10:20	AJMA	Ok
3	M8011.504.1 0.25 PPB ICC	PR060261.D	13 Mar 2023 10:30	AJMA	Ok
4	M8011.504.1 0.1 PPB ICC	PR060262.D	13 Mar 2023 10:40	AJMA	Ok
5	M8011.504.1 0.05 PPB ICC	PR060263.D	13 Mar 2023 10:49	AJMA	Ok
6	M8011.504.1 0.025 PPB ICC	PR060264.D	13 Mar 2023 10:59	AJMA	Ok,M
7	M8011.504.1 0.1 PPB ICV	PR060265.D	13 Mar 2023 11:09	AJMA	Ok
8	RT CHECK	PR060266.D	13 Mar 2023 11:19	AJMA	Ok
9	M8011-504 CCC 0.1 PPB	PR060267.D	13 Mar 2023 11:28	AJMA	Ok
10	Low-Level LFB-MDL Check	PR060268.D	13 Mar 2023 11:38	AJMA	Ok
11	PB151379BL	PR060269.D	13 Mar 2023 11:48	AJMA	Ok
12	PB151379BS	PR060270.D	13 Mar 2023 11:57	AJMA	Ok
13	PB151379BSD	PR060271.D	13 Mar 2023 12:07	AJMA	Ok
14	O1627-07	PR060272.D	13 Mar 2023 13:16	AJMA	Ok
15	O1627-08	PR060273.D	13 Mar 2023 13:26	AJMA	Ok
16	M8011-504 CCC 0.1 PPB	PR060274.D	13 Mar 2023 13:35	AJMA	Ok

M : Manual Integration



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR040623

Review By	Ankita	Review On	4/6/2023 11:53:50 AM
Supervise By	Sohil	Supervise On	4/7/2023 8:22:54 AM
SubDirectory	PR040623	HP Acquire Method	HP Processing Method PR032823
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP21869,PP21870,PP21871,PP21872,PP21873		
CCC	PP21886		
Internal Standard/PEM			
ICV/I.BLK	PP21874		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	Data File Name	Date-Time	Operator	Status
1	HEXANE	PR060683.D	06 Apr 2023 09:45	AJMA	Ok
2	M8011-504 CCC 0.1 PPB	PR060684.D	06 Apr 2023 09:55	AJMA	Ok
3	Low-Level LFB-MDL Check	PR060685.D	06 Apr 2023 10:04	AJMA	Ok
4	PB151926BL	PR060686.D	06 Apr 2023 10:16	AJMA	Ok
5	PB151926BS	PR060687.D	06 Apr 2023 10:26	AJMA	Ok
6	PB151926BSD	PR060688.D	06 Apr 2023 10:36	AJMA	Ok
7	O2213-07	PR060689.D	06 Apr 2023 11:14	AJMA	Ok
8	O2213-08	PR060690.D	06 Apr 2023 11:26	AJMA	Ok
9	M8011-504 CCC 0.1 PPB	PR060691.D	06 Apr 2023 11:38	AJMA	Ok

M : Manual Integration