



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Remington & Vernick Engineers

Project Name: 169 Bridgeton Pike Mullica Hill, NJ # 0800X062

Project # N/A

Chemtech Project # O2505

Test Name: SVOC-TCL BNA -20

A. Number of Samples and Date of Receipt:

7 Solid samples were received on 04/26/2023.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: EPH, EPH_NF, Metals Group3, PCB, SVOC-TCL BNA -20 and VOC-TCLVOA-10.

This data package contains results for SVOC-TCL BNA -20.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements except for WO-5 due to matrix interference.

The Retention Times were acceptable for all samples.

The MS {O2583-04MS} with File ID: BF133210.D recoveries met the requirements for all compounds except for 4-Chloroaniline[46%] this compound did not meet the NJDKQP criteria but met the in-house criteria also 2,3,4,6-Tetrachlorophenol[44%], 2,4,5-Trichlorophenol[65%], 2,4-Dinitrophenol[0%], 4,6-Dinitro-2-methylphenol[0%], and Pentachlorophenol[10%] these compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference.

The MSD {O2583-04MSD} with File ID: BF133211.D recoveries met the acceptable requirements except for 4-Chloroaniline[50%] this compound did not meet the NJDKQP criteria but met the in-house criteria also 2,3,4,6-Tetrachlorophenol[50%], 2,4-Dinitrophenol[0%], 4,6-Dinitro-2-methylphenol[6%], and Pentachlorophenol[15%] these compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

The RPD for {O2583-04MSD} with File ID: BF133211.D met criteria except for 4,6-Dinitro-2-methylphenol[200%], Pentachlorophenol[40%] due to difference in results of MS and MSD.

The Blank Spike for {PB152548BS} with File ID: BF133203.D met requirements for all samples except for 3-Nitroaniline[65%], 4-Chloroaniline[55%] these compounds did not meet the NJDKQP criteria but met the in-house criteria .

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration method (Method 8270-BF042123.M) for Benzaldehyde, this compound is passing on Quadratic regression.

The Continuous Calibration File ID BF133309.D met the requirements except for Hexachlorocyclopentadiene is failing marginally low therefore no corrective action taken.

The Tuning criteria met requirements.

Samples WO-4, WO-5 was diluted due to their dirty, concentrated and viscous matrix.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____