



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick Engineers

SDG No.: O2505

Contract: REMI02

Lab Code: CHEM

Case No.: O2505

SAS No.: O2505

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	1060	1000	106	90 - 110	P	05/02/2023	17:12	LB125255



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick Engineers **SDG No.:** O2505
Contract: REMI02 **Lab Code:** CHEM **Case No.:** O2505 **SAS No.:** O2505
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4870	5000	98	90 - 110	P	05/02/2023	17:31	LB125255



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick Engineers **SDG No.:** O2505
Contract: REMI02 **Lab Code:** CHEM **Case No.:** O2505 **SAS No.:** O2505
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLCCV01	Lead	13.4	12.0	112	80 - 120	P	05/02/2023	17:35	LB125255



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Remington & Vernick Engineers SDG No.: O2505
Contract: REMI02 Lab Code: CHEM Case No.: O2505 SAS No.: O2505
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Lead	4850	5000	97	90 - 110	P	05/02/2023	18:38	LB125255
CCV03	Lead	4900	5000	98	90 - 110	P	05/02/2023	19:25	LB125255
CCV04	Lead	4820	5000	96	90 - 110	P	05/02/2023	20:11	LB125255
CCV05	Lead	4800	5000	96	90 - 110	P	05/02/2023	20:57	LB125255
CCV06	Lead	4740	5000	95	90 - 110	P	05/02/2023	21:45	LB125255
CCV07	Lead	4810	5000	96	90 - 110	P	05/02/2023	22:19	LB125255