



Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Remington & Vernick Engineers **SDG No.:** O2505
Contract: REMI02 **Lab Code:** CHEM **Case No.:** O2505 **SAS No.:** O2505
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	0.23			-12	12	05/02/2023	17:24	LB125255
ICSAB01	Lead	49.1	49.0	100	37	61	05/02/2023	17:27	LB125255