



NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH QC BATCHID: LB125255

	NA	NO	YES
1.Calibration Summary met criteria.	_____	_____	_____✓_____
2.ICP Interference Check Sample Results Summary Submitted.	_____	_____	_____✓_____
3. Serial Dilution Summary (if applicable) Submitted.	_____	_____	_____✓_____
4.Laboratory Control Sample Summary (if applicable) Submitted.:	_____	_____	_____✓_____
5.Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓_____	_____
6.Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
7.Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
8.Digestion Holding Time Met If not met, list number of days exceeded for each sample:	_____	_____	_____✓_____
9.Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____

ADDITIONAL COMMENTS:

MS/MSD fail for O2495-09 (Fe,Na) parameters due to both chemicals interference during digestion and unknow sample matrix interference. Post spike still fail for Na.

O2555-01 (Sb) Parameters due to chemicals interference during digestion.

Analyst

Date _____

QA REVIEW

Date _____