



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB125255

Review By	bin	Review On	5/3/2023 10:12:19 AM
Supervise By	jaswal	Supervise On	5/4/2023 3:14:45 AM

STD. NAME	STD REF.#
ICAL Standard	MP75083,MP75140,MP75141,MP75142,MP75143,MP75148
ICV Standard	MP75151,MP75148
CCV Standard	MP75144
ICSA Standard	MP75138,MP75139
CRI Standard	
LCS Standard	
Chk Standard	MP75137,MP75152,MP75149,MP75150

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	05/02/23 16:27		Bin	OK
2	S1	S1	CAL2	05/02/23 16:31		Bin	OK
3	S2	S2	CAL3	05/02/23 16:35		Bin	OK
4	S3	S3	CAL4	05/02/23 16:39		Bin	OK
5	S4	S4	CAL5	05/02/23 16:43		Bin	OK
6	S5	S5	CAL6	05/02/23 16:47		Bin	OK
7	ICV01	ICV01	ICV	05/02/23 17:12		Bin	OK
8	ICB01	ICB01	ICB	05/02/23 17:16		Bin	OK
9	CRI01	CRI01	CRDL	05/02/23 17:20		Bin	OK
10	ICSA01	ICSA01	ICSA	05/02/23 17:24		Bin	OK
11	ICSAB01	ICSAB01	ICSAB	05/02/23 17:27		Bin	OK
12	CCV01	CCV01	CCV	05/02/23 17:31		Bin	OK
13	LLCCV01	LLCCV01	LLCCV	05/02/23 17:35		Bin	OK
14	CCB01	CCB01	CCB	05/02/23 17:39		Bin	OK
15	O2550-05DL	M-041923RB09DL	SAM	05/02/23 17:43	10X For Ag,Se	Bin	Confirms
16	O2416-05	4-GRAB	SAM	05/02/23 17:51		Bin	OK
17	PB152420BL	PB152420BL	MB	05/02/23 17:59		Bin	OK
18	PB152420BS	PB152420BS	LCS	05/02/23 18:03		Bin	OK
19	O2495-09	GMW-5	SAM	05/02/23 18:07		Bin	OK
20	O2495-09DUP	GMW-5DUP	DUP	05/02/23 18:11		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB125255

Review By	bin	Review On	5/3/2023 10:12:19 AM				
Supervise By	jaswal	Supervise On	5/4/2023 3:14:45 AM				
STD. NAME		STD REF.#					
ICAL Standard		MP75083,MP75140,MP75141,MP75142,MP75143,MP75148					
ICV Standard		MP75151,MP75148					
CCV Standard		MP75144					
ICSA Standard		MP75138,MP75139					
CRI Standard							
LCS Standard							
Chk Standard		MP75137,MP75152,MP75149,MP75150					

21	O2416-03	2-GRAB	SAM	05/02/23 18:15		Bin	OK
22	LR CHECK 1	LR CHECK 1	HIGH STD	05/02/23 18:19		Bin	OK
23	O2389-07DL	SPB-2WC-23-04-17D	SAM	05/02/23 18:24	100X For Zn	Bin	Confirm
24	LR CHECK 2	LR CHECK 2	HIGH STD	05/02/23 18:28		Bin	OK
25	CCV02	CCV02	CCV	05/02/23 18:38		Bin	OK
26	CCB02	CCB02	CCB	05/02/23 18:42		Bin	OK
27	O2495-09L	GMW-5L	SD	05/02/23 18:47		Bin	OK
28	O2495-09MS	GMW-5MS	MS	05/02/23 18:51		Bin	OK
29	O2495-09MSD	GMW-5MSD	MSD	05/02/23 18:55		Bin	OK
30	O2495-09A	GMW-5A	PS	05/02/23 18:59	0.1 ml of LFS-1(M5318) and LF	Bin	OK
31	PB152536BL	PB152536BL	MB	05/02/23 19:02		Bin	OK
32	PB152536BS	PB152536BS	LCS	05/02/23 19:06		Bin	OK
33	O2505-04	WO-4	SAM	05/02/23 19:10		Bin	OK
34	O2505-05	WO-5	SAM	05/02/23 19:14		Bin	OK
35	O2555-01	OE-1-042823	SAM	05/02/23 19:18		Bin	OK
36	O2555-01DUP	OE-1-042823DUP	DUP	05/02/23 19:22		Bin	OK
37	CCV03	CCV03	CCV	05/02/23 19:25		Bin	OK
38	CCB03	CCB03	CCB	05/02/23 19:29		Bin	OK
39	O2555-01L	OE-1-042823L	SD	05/02/23 19:33		Bin	OK
40	O2555-01MS	OE-1-042823MS	MS	05/02/23 19:37		Bin	OK
41	O2555-01MSD	OE-1-042823MSD	MSD	05/02/23 19:41		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB125255

Review By		bin		Review On		5/3/2023 10:12:19 AM	
Supervise By		jaswal		Supervise On		5/4/2023 3:14:45 AM	
STD. NAME		STD REF.#					
ICAL Standard		MP75083,MP75140,MP75141,MP75142,MP75143,MP75148					
ICV Standard		MP75151,MP75148					
CCV Standard		MP75144					
ICSA Standard		MP75138,MP75139					
CRI Standard							
LCS Standard							
Chk Standard		MP75137,MP75152,MP75149,MP75150					

42	O2555-01A	OE-1-042823A	PS	05/02/23 19:44	0.1 ml of LFS-1(M5318) and LF	Bin	OK
43	O2557-01	JC-03-042823-A	SAM	05/02/23 19:48		Bin	OK
44	O2557-03	JC-03-042823-B	SAM	05/02/23 19:52		Bin	OK
45	O2562-01	OR-02-04282023	SAM	05/02/23 19:56		Bin	OK
46	O2563-01	HR-01-042823	SAM	05/02/23 19:59		Bin	OK
47	O2563-03	HR-03-042823	SAM	05/02/23 20:03		Bin	OK
48	O2567-01	B-21-SB01	SAM	05/02/23 20:07	Cu high	Bin	Dilution
49	CCV04	CCV04	CCV	05/02/23 20:11		Bin	OK
50	CCB04	CCB04	CCB	05/02/23 20:15		Bin	OK
51	O2567-02	B-21-SB02	SAM	05/02/23 20:19	CCV Fail For V	Bin	Not Ok
52	O2568-01	72-11938	SAM	05/02/23 20:22	CCV Fail For V	Bin	Not Ok
53	O2569-01	ETGI-339	SAM	05/02/23 20:26	CCV Fail For V	Bin	Not Ok
54	O2570-01	RB-1	SAM	05/02/23 20:30	CCV Fail For V	Bin	Not Ok
55	O2571-01	OE-3-050123	SAM	05/02/23 20:34	CCV Fail For V	Bin	Not Ok
56	O2572-01	RBR-251671	SAM	05/02/23 20:38	CCV Fail For V	Bin	Not Ok
57	O2572-03	RT2343	SAM	05/02/23 20:41	CCV Fail For V	Bin	Not Ok
58	O2495-01	PZ-01	SAM	05/02/23 20:45	CCV Fail For V	Bin	Not Ok
59	O2495-02	PZ-01	SAM	05/02/23 20:49	CCV Fail For V	Bin	Not Ok
60	O2495-03	PZ-02	SAM	05/02/23 20:53	CCV Fail For V	Bin	Not Ok
61	CCV05	CCV05	CCV	05/02/23 20:57	CCV Fail For V	Bin	OK
62	CCB05	CCB05	CCB	05/02/23 21:01		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB125255

Review By	bin	Review On	5/3/2023 10:12:19 AM
Supervise By	jaswal	Supervise On	5/4/2023 3:14:45 AM

STD. NAME	STD REF.#
ICAL Standard	MP75083,MP75140,MP75141,MP75142,MP75143,MP75148
ICV Standard	MP75151,MP75148
CCV Standard	MP75144
ICSA Standard	MP75138,MP75139
CRI Standard	
LCS Standard	
Chk Standard	MP75137,MP75152,MP75149,MP75150

63	O2495-04	PZ-02	SAM	05/02/23 21:05	CCV Fail For Ba,V	Bin	Not Ok
64	O2495-05	GMW-11	SAM	05/02/23 21:09	CCV Fail For Ba,V	Bin	Not Ok
65	O2495-06	GMW-11	SAM	05/02/23 21:13	CCV Fail For Ba,V	Bin	Not Ok
66	O2495-07	GMW-X	SAM	05/02/23 21:17	CCV Fail For Ba,V	Bin	Not Ok
67	O2495-08	GMW-X	SAM	05/02/23 21:21	CCV Fail For Ba,V	Bin	Not Ok
68	O2495-10	GMW-5	SAM	05/02/23 21:25	CCV Fail For Ba,V	Bin	Not Ok
69	O2492-01	22-MW-04I	SAM	05/02/23 21:29		Bin	OK
70	O2492-02	22-MW-04I	SAM	05/02/23 21:33		Bin	OK
71	O2492-03	LEA-5SR	SAM	05/02/23 21:37		Bin	OK
72	O2492-04	LEA-5SR	SAM	05/02/23 21:41		Bin	OK
73	CCV06	CCV06	CCV	05/02/23 21:45	CCV Fail For Ba,V	Bin	OK
74	CCB06	CCB06	CCB	05/02/23 21:49		Bin	OK
75	O2492-05	LEA-8I	SAM	05/02/23 21:53		Bin	OK
76	O2492-06	LEA-8I	SAM	05/02/23 21:57		Bin	OK
77	O2535-03	TP-8	SAM	05/02/23 22:01		Bin	OK
78	O2535-04	TP-8	SAM	05/02/23 22:06		Bin	OK
79	O2535-05	TP-7	SAM	05/02/23 22:10		Bin	OK
80	O2535-06	TP-7	SAM	05/02/23 22:14		Bin	OK
81	CCV07	CCV07	CCV	05/02/23 22:19		Bin	OK
82	CCB07	CCB07	CCB	05/02/23 22:22		Bin	OK