



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: FID_D

Daily Analysis Runlog For Sequence/QC Batch ID # FD050423AR

Review By	yogesh	Review On	5/4/2023 9:59:48 AM				
Supervise By	Ankita	Supervise On	5/5/2023 9:28:15 AM				
SubDirectory	FD050423AR	HP Acquire Method	HP Processing Method		FD042123AR		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		PP21759,PP21858,PP21859,PP21860,PP21861					
		PP21762,PP21859					
		PP21760,PP21862					
Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FD045578.D	04 May 2023 07:59		YP/AJ	Ok
2	I.BLK	I.BLK	FD045579.D	04 May 2023 08:36		YP/AJ	Ok
3	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FD045580.D	04 May 2023 09:15		YP/AJ	Ok
4	PB152543BL	PB152543BL	FD045581.D	04 May 2023 09:53		YP/AJ	Ok
5	O2500-17	PT-NJEPH-SOIL	FD045582.D	04 May 2023 10:32	need 5x&10X dilution	YP/AJ	Dilution
6	O2500-17DL	PT-NJEPH-SOILDL	FD045583.D	04 May 2023 11:19	need further dilution	YP/AJ	Dilution
7	O2500-17DL2	PT-NJEPH-SOILDL2	FD045584.D	04 May 2023 11:57		YP/AJ	Ok,M
8	I.BLK	I.BLK	FD045585.D	04 May 2023 12:35		YP/AJ	Ok
9	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FD045586.D	04 May 2023 13:52		YP/AJ	Ok
10	PB152571BL	PB152571BL	FD045587.D	04 May 2023 15:13		YP/AJ	Ok
11	PB152571BS	PB152571BS	FD045588.D	04 May 2023 15:51		YP/AJ	Ok
12	PB152571BSD	PB152571BSD	FD045589.D	04 May 2023 16:30		YP/AJ	Ok
13	O2567-03	EB-2023-29-06	FD045590.D	04 May 2023 17:09		YP/AJ	Ok
14	O2505-08DL	WO-5DL	FD045591.D	04 May 2023 17:47	need further dilution	YP/AJ	Dilution
15	O2505-08DL2	WO-5DL2	FD045592.D	04 May 2023 18:25		YP/AJ	Ok
16	I.BLK	I.BLK	FD045593.D	04 May 2023 19:04		YP/AJ	Ok
17	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FD045594.D	04 May 2023 19:42		YP/AJ	Ok

M : Manual Integration