



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050323AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063531.D

Aliphatic C9-C12	8623493.000	60.000	3.197	6.484	143724.883	132318.329	-8.621
Aliphatic C12-C16	5770579.000	40.000	6.485	9.878	144264.475	130577.891	-10.482
Aliphatic C16-C21	8672293.000	60.000	9.879	13.237	144538.217	132388.089	-9.178
Aliphatic C21-C28	11162560.000	80.000	13.238	16.890	139532.000	128989.378	-8.173
Aliphatic C28-C40	14693599.000	120.000	16.891	21.743	122446.658	122159.118	-0.235
Aliphatic EPH	48922524.000	360.000	3.197	21.743	135895.900	128010.403	-6.160