



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050323AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063542.D

Aliphatic C9-C12	8918811.000	60.000	3.197	6.484	148646.850	132318.329	-12.340
Aliphatic C12-C16	5955250.000	40.000	6.485	9.878	148881.250	130577.891	-14.017
Aliphatic C16-C21	9017986.000	60.000	9.879	13.237	150299.767	132388.089	-13.530
Aliphatic C21-C28	11538453.000	80.000	13.238	16.890	144230.663	128989.378	-11.816
Aliphatic C28-C40	15700301.000	120.000	16.891	21.743	130835.842	122159.118	-7.103
Aliphatic EPH	51130801.000	360.000	3.197	21.743	142030.003	128010.403	-10.952