



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050423AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063545.D

Aliphatic C9-C12	8813581.000	60.000	3.197	6.482	146893.017	132318.329	-11.015
Aliphatic C12-C16	5868566.000	40.000	6.483	9.875	146714.150	130577.891	-12.358
Aliphatic C16-C21	8817012.000	60.000	9.876	13.231	146950.200	132388.089	-11.000
Aliphatic C21-C28	11402684.000	80.000	13.232	16.883	142533.550	128989.378	-10.500
Aliphatic C28-C40	14911032.000	120.000	16.884	21.732	124258.600	122159.118	-1.719
Aliphatic EPH	49812875.000	360.000	3.197	21.732	138369.097	128010.403	-8.092