



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050423AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063552.D

Aliphatic C9-C12	8442959.000	60.000	3.197	6.482	140715.983	132318.329	-6.347
Aliphatic C12-C16	5670965.000	40.000	6.483	9.875	141774.125	130577.891	-8.574
Aliphatic C16-C21	8588817.000	60.000	9.876	13.231	143146.950	132388.089	-8.127
Aliphatic C21-C28	11032596.000	80.000	13.232	16.883	137907.450	128989.378	-6.914
Aliphatic C28-C40	14531231.000	120.000	16.884	21.732	121093.592	122159.118	0.872
Aliphatic EPH	48266568.000	360.000	3.197	21.732	134073.800	128010.403	-4.737