



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050423AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063560.D

Aliphatic C9-C12	8485228.000	60.000	3.197	6.482	141420.467	132318.329	-6.879
Aliphatic C12-C16	5664089.000	40.000	6.483	9.875	141602.225	130577.891	-8.443
Aliphatic C16-C21	8521672.000	60.000	9.876	13.231	142027.867	132388.089	-7.281
Aliphatic C21-C28	11010838.000	80.000	13.232	16.883	137635.475	128989.378	-6.703
Aliphatic C28-C40	15059816.000	120.000	16.884	21.732	125498.467	122159.118	-2.734
Aliphatic EPH	48741643.000	360.000	3.197	21.732	135393.453	128010.403	-5.768