



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD050323AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD045566.D

Aromatic C10-C12	7416122.000	40.000	4.249	5.981	185403.050	169348.262	-9.480
Aromatic C12-C16	10560916.000	60.000	5.982	8.598	176015.267	160865.134	-9.418
Aromatic C16-C21	12481103.000	80.000	8.599	12.867	156013.788	143957.476	-8.375
Aromatic C21-C36	25592079.000	180.000	12.868	18.259	142178.217	133648.613	-6.382
Aromatic EPH	56050220.000	360.000	4.249	18.259	155695.056	144442.186	-7.791