



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD050423AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD045580.D

Aromatic C10-C12	7354415.000	40.000	4.253	5.985	183860.375	169348.262	-8.569
Aromatic C12-C16	10499024.000	60.000	5.986	8.601	174983.733	160865.134	-8.777
Aromatic C16-C21	12620458.000	80.000	8.602	12.869	157755.725	143957.476	-9.585
Aromatic C21-C36	24154795.000	180.000	12.870	18.259	134193.306	133648.613	-0.408
Aromatic EPH	54628692.000	360.000	4.253	18.259	151746.367	144442.186	-5.057