



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD050423AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD045586.D

Aromatic C10-C12	6785536.000	40.000	4.253	5.985	169638.400	169348.262	-0.171
Aromatic C12-C16	9722755.000	60.000	5.986	8.601	162045.917	160865.134	-0.734
Aromatic C16-C21	11489853.000	80.000	8.602	12.869	143623.163	143957.476	0.232
Aromatic C21-C36	23205389.000	180.000	12.870	18.259	128918.828	133648.613	3.539
Aromatic EPH	51203533.000	360.000	4.253	18.259	142232.036	144442.186	1.530