



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD050423AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD045594.D

Aromatic C10-C12	7543036.000	40.000	4.253	5.985	188575.900	169348.262	-11.354
Aromatic C12-C16	10844903.000	60.000	5.986	8.601	180748.383	160865.134	-12.360
Aromatic C16-C21	13155133.000	80.000	8.602	12.869	164439.163	143957.476	-14.228
Aromatic C21-C36	25312131.000	180.000	12.870	18.259	140622.950	133648.613	-5.218
Aromatic EPH	56855203.000	360.000	4.253	18.259	157931.119	144442.186	-9.339