



SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name:

Chemtech

Client:

Remington & Vernick Engineers

Lab Code:

CHEM

Cas No:

O2505

SAS No :

O2505

SDG No:

O2505

Sample No :

PB152543BS

Datafile:

FC063533.D

Client ID :

PB152543BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10	7.05	70		(40-140)
Aliphatic C12-C16	6.7	6.88	103		(40-140)
Aliphatic C16-C21	10	11.6	116		(40-140)
Aliphatic C21-C28	13.3	15.5	116		(40-140)
Aliphatic C28-C40	20.0	19.6	98		(40-140)