



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC042723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063485.D

Aliphatic C9-C12	8276929.000	60.000	3.199	6.486	137948.817	137065.155	-0.645
Aliphatic C12-C16	5437066.000	40.000	6.487	9.878	135926.650	135585.936	-0.251
Aliphatic C16-C21	8322917.000	60.000	9.879	13.234	138715.283	135817.180	-2.134
Aliphatic C21-C28	11199468.000	80.000	13.235	16.884	139993.350	134266.069	-4.266
Aliphatic C28-C40	17065847.000	120.000	16.885	21.741	142215.392	131023.808	-8.542
Aliphatic EPH	50302227.000	360.000	3.199	21.741	139728.408	134057.000	-4.231