



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE042723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE041608.D

Aliphatic C9-C12	9282297.000	60.000	3.256	6.877	154704.950	149673.029	-3.362
Aliphatic C12-C16	6417165.000	40.000	6.878	10.315	160429.125	155744.024	-3.008
Aliphatic C16-C21	9485191.000	60.000	10.316	13.682	158086.517	161376.097	2.038
Aliphatic C21-C28	11695093.000	80.000	13.683	17.343	146188.663	158129.852	7.552
Aliphatic C28-C40	16931388.000	120.000	17.344	22.280	141094.900	139105.719	-1.430
Aliphatic EPH	53811134.000	360.000	3.256	22.280	149475.372	150654.953	0.783