



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE050123AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE041623.D

Aliphatic C9-C12	10340344.000	60.000	3.258	6.881	172339.067	149673.029	-15.144
Aliphatic C12-C16	7126803.000	40.000	6.882	10.321	178170.075	155744.024	-14.399
Aliphatic C16-C21	10557201.000	60.000	10.322	13.691	175953.350	161376.097	-9.033
Aliphatic C21-C28	13010510.000	80.000	13.692	17.357	162631.375	158129.852	-2.847
Aliphatic C28-C40	17698986.000	120.000	17.358	22.307	147491.550	139105.719	-6.028
Aliphatic EPH	58733844.000	360.000	3.258	22.307	163149.567	150654.953	-8.294