



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE050123AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE041630.D

Aliphatic C9-C12	9822775.000	60.000	3.258	6.881	163712.917	149673.029	-9.380
Aliphatic C12-C16	6766959.000	40.000	6.882	10.321	169173.975	155744.024	-8.623
Aliphatic C16-C21	10137774.000	60.000	10.322	13.691	168962.900	161376.097	-4.701
Aliphatic C21-C28	12579576.000	80.000	13.692	17.357	157244.700	158129.852	0.560
Aliphatic C28-C40	17570696.000	120.000	17.358	22.307	146422.467	139105.719	-5.260
Aliphatic EPH	56877780.000	360.000	3.258	22.307	157993.833	150654.953	-4.871