



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE050123AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE041652.D

Aliphatic C9-C12	8955589.000	60.000	3.258	6.881	149259.817	149673.029	0.276
Aliphatic C12-C16	6205537.000	40.000	6.882	10.321	155138.425	155744.024	0.389
Aliphatic C16-C21	9283228.000	60.000	10.322	13.691	154720.467	161376.097	4.124
Aliphatic C21-C28	11598080.000	80.000	13.692	17.357	144976.000	158129.852	8.318
Aliphatic C28-C40	16457146.000	120.000	17.358	22.307	137142.883	139105.719	1.411
Aliphatic EPH	52499580.000	360.000	3.258	22.307	145832.167	150654.953	3.201