



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE050123AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE041660.D

Aliphatic C9-C12	8986490.000	60.000	3.258	6.881	149774.833	149673.029	-0.068
Aliphatic C12-C16	6255217.000	40.000	6.882	10.321	156380.425	155744.024	-0.409
Aliphatic C16-C21	9392091.000	60.000	10.322	13.691	156534.850	161376.097	3.000
Aliphatic C21-C28	11776816.000	80.000	13.692	17.357	147210.200	158129.852	6.905
Aliphatic C28-C40	16636043.000	120.000	17.358	22.307	138633.692	139105.719	0.339
Aliphatic EPH	53046657.000	360.000	3.258	22.307	147351.825	150654.953	2.193