



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 12:34 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1254-1 (1)   | 6.46       | 6.46      | 6.36      | 6.56  | 0.00       |
| Aroclor-1254-2 (2)   | 6.68       | 6.68      | 6.58      | 6.78  | 0.00       |
| Aroclor-1254-3 (3)   | 7.05       | 7.05      | 6.95      | 7.15  | 0.00       |
| Aroclor-1254-4 (4)   | 7.34       | 7.34      | 7.24      | 7.44  | 0.00       |
| Aroclor-1254-5 (5)   | 7.77       | 7.76      | 7.66      | 7.86  | -0.01      |
| Tetrachloro-m-xylene | 4.42       | 4.42      | 4.32      | 4.52  | 0.00       |
| Decachlorobiphenyl   | 10.27      | 10.25     | 10.15     | 10.35 | -0.02      |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 12:34 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.50       | 5.50      | 5.40      | 5.60 | 0.00       |
| Aroclor-1254-2 (2)   | 5.65       | 5.65      | 5.55      | 5.75 | 0.00       |
| Aroclor-1254-3 (3)   | 6.05       | 6.05      | 5.95      | 6.15 | 0.00       |
| Aroclor-1254-4 (4)   | 6.28       | 6.28      | 6.18      | 6.38 | 0.00       |
| Aroclor-1254-5 (5)   | 6.70       | 6.70      | 6.60      | 6.80 | 0.00       |
| Tetrachloro-m-xylene | 3.62       | 3.62      | 3.52      | 3.72 | 0.00       |
| Decachlorobiphenyl   | 8.64       | 8.63      | 8.53      | 8.73 | -0.01      |