



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 16:30 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.60	5.60	5.50	5.70	0.00
Aroclor-1242-2 (2)	5.62	5.62	5.52	5.72	0.00
Aroclor-1242-3 (3)	5.68	5.68	5.58	5.78	0.00
Aroclor-1242-4 (4)	5.78	5.78	5.68	5.88	0.00
Aroclor-1242-5 (5)	6.53	6.52	6.42	6.62	-0.01
Tetrachloro-m-xylene	4.42	4.42	4.32	4.52	0.00
Decachlorobiphenyl	10.27	10.26	10.16	10.36	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 16:30 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.70	4.70	4.60	4.80	0.00
Aroclor-1242-2 (2)	4.72	4.72	4.62	4.82	0.00
Aroclor-1242-3 (3)	4.89	4.90	4.80	5.00	0.01
Aroclor-1242-4 (4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1242-5 (5)	5.50	5.50	5.40	5.60	0.00
Tetrachloro-m-xylene	3.62	3.62	3.52	3.72	0.00
Decachlorobiphenyl	8.64	8.63	8.53	8.73	-0.01