



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 21:41 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.60	5.59	5.49	5.69	0.00
Aroclor-1016-2	(2)	5.62	5.61	5.51	5.71	-0.01
Aroclor-1016-3	(3)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-4	(4)	5.78	5.78	5.68	5.88	0.00
Aroclor-1016-5	(5)	6.08	6.07	5.97	6.17	-0.01
Aroclor-1260-1	(1)	7.22	7.21	7.11	7.31	-0.01
Aroclor-1260-2	(2)	7.48	7.47	7.37	7.57	-0.01
Aroclor-1260-3	(3)	7.84	7.84	7.74	7.94	0.00
Aroclor-1260-4	(4)	8.07	8.06	7.96	8.16	-0.01
Aroclor-1260-5	(5)	8.40	8.39	8.29	8.49	-0.01
Tetrachloro-m-xylene		4.42	4.41	4.31	4.51	-0.01
Decachlorobiphenyl		10.26	10.25	10.15	10.35	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 21:41 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.70	4.70	4.60	4.80	0.00
Aroclor-1016-2	(2)	4.72	4.72	4.62	4.82	0.00
Aroclor-1016-3	(3)	4.89	4.89	4.79	4.99	0.00
Aroclor-1016-4	(4)	4.94	4.94	4.84	5.04	0.01
Aroclor-1016-5	(5)	5.15	5.15	5.05	5.25	0.00
Aroclor-1260-1	(1)	6.18	6.18	6.08	6.28	0.00
Aroclor-1260-2	(2)	6.37	6.37	6.27	6.47	0.00
Aroclor-1260-3	(3)	6.53	6.53	6.43	6.63	0.00
Aroclor-1260-4	(4)	7.00	7.00	6.90	7.10	0.00
Aroclor-1260-5	(5)	7.24	7.24	7.14	7.34	0.00
Tetrachloro-m-xylene		3.62	3.62	3.52	3.72	0.00
Decachlorobiphenyl		8.63	8.63	8.53	8.73	0.00