



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 22:15 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1248-1	(1)	5.59	5.59	5.49	5.69	0.00
Aroclor-1248-2	(2)	5.87	5.87	5.77	5.97	0.00
Aroclor-1248-3	(3)	6.08	6.08	5.98	6.18	0.00
Aroclor-1248-4	(4)	6.49	6.49	6.39	6.59	0.00
Aroclor-1248-5	(5)	6.52	6.52	6.42	6.62	0.00
Tetrachloro-m-xylene		4.42	4.42	4.32	4.52	0.01
Decachlorobiphenyl		10.26	10.26	10.16	10.36	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 22:15 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1248-1 (1)	4.70	4.70	4.60	4.80	0.00
Aroclor-1248-2 (2)	4.94	4.94	4.84	5.04	0.01
Aroclor-1248-3 (3)	4.98	4.98	4.88	5.08	0.00
Aroclor-1248-4 (4)	5.15	5.15	5.05	5.25	0.00
Aroclor-1248-5 (5)	5.54	5.54	5.44	5.64	0.00
Tetrachloro-m-xylene	3.62	3.62	3.52	3.72	0.00
Decachlorobiphenyl	8.63	8.63	8.53	8.73	0.00