



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 22:32 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1254-1	(1)	6.46	6.46	6.36	6.56	0.00
Aroclor-1254-2	(2)	6.68	6.68	6.58	6.78	0.00
Aroclor-1254-3	(3)	7.05	7.05	6.95	7.15	0.00
Aroclor-1254-4	(4)	7.34	7.34	7.24	7.44	0.00
Aroclor-1254-5	(5)	7.76	7.76	7.66	7.86	0.00
Tetrachloro-m-xylene		4.42	4.42	4.32	4.52	0.00
Decachlorobiphenyl		10.26	10.25	10.15	10.35	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/03/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 22:32 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.50	5.50	5.40	5.60	0.00
Aroclor-1254-2 (2)	5.65	5.65	5.55	5.75	0.00
Aroclor-1254-3 (3)	6.05	6.05	5.95	6.15	0.00
Aroclor-1254-4 (4)	6.28	6.28	6.18	6.38	0.00
Aroclor-1254-5 (5)	6.70	6.70	6.60	6.80	0.00
Tetrachloro-m-xylene	3.62	3.62	3.52	3.72	0.00
Decachlorobiphenyl	8.63	8.63	8.53	8.73	0.00