



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/04/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 15:52 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1248-1       | (1) | 5.59       | 5.59      | 5.49                   | 5.69  | 0.00       |
| Aroclor-1248-2       | (2) | 5.87       | 5.87      | 5.77                   | 5.97  | 0.00       |
| Aroclor-1248-3       | (3) | 6.08       | 6.08      | 5.98                   | 6.18  | 0.00       |
| Aroclor-1248-4       | (4) | 6.48       | 6.49      | 6.39                   | 6.59  | 0.01       |
| Aroclor-1248-5       | (5) | 6.52       | 6.52      | 6.42                   | 6.62  | 0.00       |
| Tetrachloro-m-xylene |     | 4.42       | 4.42      | 4.32                   | 4.52  | 0.01       |
| Decachlorobiphenyl   |     | 10.26      | 10.26     | 10.16                  | 10.36 | 0.00       |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: REMI02

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

Continuing Calib Date: 05/04/2023 Initial Calibration Date(s): 04/26/2023 04/26/2023

Continuing Calib Time: 15:52 Initial Calibration Time(s): 15:39 23:18

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Aroclor-1248-1 (1)   | 4.72       | 4.70      | 4.60                   | 4.80 | -0.02      |
| Aroclor-1248-4 (4)   | 5.15       | 5.15      | 5.05                   | 5.25 | 0.00       |
| Tetrachloro-m-xylene | 1.12       | 3.62      | 3.52                   | 3.72 | 2.50       |
| Decachlorobiphenyl   | 6.13       | 8.63      | 8.53                   | 8.73 | 2.50       |