



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: O2505

Client: Remington & Vernick Engineers

Analytical Method: 8082A

Datafile : PO094691.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB152545BS	AR1016	166.5	143	ug/kg	86				40 (71)	140 (120)	
	AR1260	166.5	131	ug/kg	79				40 (65)	140 (130)	