



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: 02505 SAS No.: 02505 SDG NO.: 02505

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/03/2023

Lab File ID: BF133200.D Time Analyzed: 11:25

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	197648	6.734	758638	8.02	395162	9.77
UPPER LIMIT	395296	7.234	1517280	8.516	790324	10.269
LOWER LIMIT	98824	6.234	379319	7.516	197581	9.269
EPA SAMPLE NO.						
01 CF-24MS	197905	6.73	818790	8.02	422731	9.77
02 CF-24MSD	177861	6.73	740021	8.02	377177	9.77
03 PB152548BL	231663	6.73	994899	8.02	533275	9.77
04 PB152548BS	194406	6.73	772245	8.02	416279	9.77

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.