



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/09/2023

Lab File ID: BF133309.D Time Analyzed: 10:04

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	214901	6.722	819732	8.01	422663	9.76
UPPER LIMIT	429802	7.222	1639460	8.51	845326	10.263
LOWER LIMIT	107451	6.222	409866	7.51	211332	9.263
EPA SAMPLE NO.						
01 WO-4	206523	6.72	780947	8.00	355273	9.76
02 WO-5	202754	6.72	750799	8.00	323955	9.76

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.