

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
 Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505
 EPA Sample No.: SSTDCCC040 Date Analyzed: 05/03/2023
 Lab File ID: BF133200.D Time Analyzed: 11:25
 Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	197648	6.734	758638	8.02	395162	9.77
UPPER LIMIT	395296	7.234	1517280	8.516	790324	10.2690
LOWER LIMIT	98824	6.234	379319	7.516	197581	9.26900
EPA SAMPLE NO.						
01 PB152548BS	194406	6.73	772245	8.02	416279	9.77
02 CF-24MS	197905	6.73	818790	8.02	422731	9.77
03 CF-24MSD	177861	6.73	740021	8.02	377177	9.77
04 PB152548BL	231663	6.73	994899	8.02	533275	9.77

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/09/2023

Lab File ID: BF133309.D Time Analyzed: 10:04

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	214901	6.722	819732	8.01	422663	9.76
UPPER LIMIT	429802	7.222	1639460	8.51	845326	10.2629
LOWER LIMIT	107451	6.222	409866	7.51	211332	9.26299
EPA SAMPLE NO.						
01 WO-4	206523	6.72	780947	8.00	355273	9.76
02 WO-5	202754	6.72	750799	8.00	323955	9.76

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.