



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: REMI02
Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505
Lab File ID: VX035577.D Date Analyzed: 05/10/2023
Instrument ID: MSVOA_X Time Analyzed: 11:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	304037	5.55	472336	6.76	419791	10.06
UPPER LIMIT	608074	6.05	944672	7.257	839582	10.555
LOWER LIMIT	152019	5.05	236168	6.257	209896	9.555
EPA SAMPLE NO.						
WO-5ME	320122	5.54	523267	6.76	483217	10.06
VX0510MBL01	303684	5.54	486455	6.76	448788	10.06
VX0510MBS01	332854	5.55	511449	6.76	454101	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.