



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: REMI02
Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505
Lab File ID: VY013569.D Date Analyzed: 05/03/2023
Instrument ID: MSVOA_Y Time Analyzed: 09:38
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	278513	7.79	413285	8.69	371747	11.49
UPPER LIMIT	557026	8.289	826570	9.191	743494	11.99
LOWER LIMIT	139257	7.289	206643	8.191	185874	10.99
EPA SAMPLE NO.						
WO-4	176608	7.79	265223	8.69	172034 *	11.49
VY0503SBL01	232503	7.80	377616	8.69	341722	11.49
VY0503SBS01	271635	7.79	413118	8.69	374754	11.49
VY0503SBSD01	284383	7.79	425765	8.69	385661	11.49

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.