



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: REMI02
Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505
Lab File ID: VX035577.D Date Analyzed: 05/10/2023
Instrument ID: MSVOA_X Time Analyzed: 11:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	209220	12.018				
UPPER LIMIT	418440	12.518				
LOWER LIMIT	104610	11.518				
EPA SAMPLE NO.						
WO-5ME	242563	12.02				
VX0510MBL01	216205	12.02				
VX0510MBS01	232176	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.