



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: REMI02
Lab Code: CHEM Case No.: O2505 SAS No.: O2505 SDG NO.: O2505
Lab File ID: VY013593.D Date Analyzed: 05/04/2023
Instrument ID: MSVOA_Y Time Analyzed: 10:51
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	170159	13.428				
UPPER LIMIT	340318	13.928				
LOWER LIMIT	85079.5	12.928				
EPA SAMPLE NO.						
WO-4RE	51026 *	13.42				
WO-5	38455 *	13.43				
VY0504SBL01	176412	13.42				
VY0504SBS01	192181	13.43				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.