



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX051023

Review By	John Carlone	Review On	5/11/2023 10:02:29 AM				
Supervise By	Krupa Patel	Supervise On	5/11/2023 12:16:17 PM				
SubDirectory	VX051023	HP Acquire Method	HP Processing Method		82X042523W.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP120449					
		VP120450,VP120451					
Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX035576.D	10 May 2023 09:41		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX035577.D	10 May 2023 11:03	pH#Lot#V12668	JC/MD	Ok,M
3	VX0510MBL01	VX0510MBL01	VX035578.D	10 May 2023 11:39		JC/MD	Ok
4	VX0510WBL01	VX0510WBL01	VX035579.D	10 May 2023 12:02		JC/MD	Ok
5	VX0510MBS01	VX0510MBS01	VX035580.D	10 May 2023 12:30		JC/MD	Ok,M
6	VX0510MBSD01	VX0510MBSD01	VX035581.D	10 May 2023 13:07		JC/MD	Ok,M
7	O2666-01	BU-02-050823	VX035582.D	10 May 2023 13:30		JC/MD	Ok
8	O2505-05ME	WO-5ME	VX035583.D	10 May 2023 13:53	Methyl acetate hit is not in first run	JC/MD	Ok
9	VX0510WBS01	VX0510WBS01	VX035584.D	10 May 2023 14:17		JC/MD	Ok,M
10	VX0510WBSD01	VX0510WBSD01	VX035585.D	10 May 2023 14:50		JC/MD	Ok,M
11	O2658-05	507	VX035586.D	10 May 2023 15:14	vial A pH<2	JC/MD	Ok
12	O2648-03DL	ORA-1636DL	VX035587.D	10 May 2023 15:37	vial B pH<2	JC/MD	Ok
13	VX0510MBSD01	VX0510MBSD01	VX035588.D	10 May 2023 16:00		JC/MD	Ok,M
14	O2658-07DL	509DL	VX035589.D	10 May 2023 16:23	vial B pH<2	JC/MD	Ok,M
15	O2719-14	TB-01_20230509	VX035590.D	10 May 2023 16:46	vial A pH<2 TB, Hit of Comp. #49	JC/MD	ReRun
16	O2723-12	TB-01-030723	VX035591.D	10 May 2023 17:10	vial A pH<2 TB	JC/MD	Ok
17	O2719-12	FB-01_20230509	VX035592.D	10 May 2023 17:33	vial A pH<2	JC/MD	Ok
18	O2746-03	TB-01-051023	VX035593.D	10 May 2023 17:56	vial A pH<2 TB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX051023

Review By	John Carlone	Review On	5/11/2023 10:02:29 AM				
Supervise By	Krupa Patel	Supervise On	5/11/2023 12:16:17 PM				
SubDirectory	VX051023	HP Acquire Method	HP Processing Method		82X042523W.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		VP120449					
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP120450,VP120451					
19	O2743-07	DRUM-A-WATER	VX035594.D	10 May 2023 18:19	vial A pH<2	JC/MD	Ok
20	O2744-01	PL-OILYWATER	VX035595.D	10 May 2023 18:43	vial A pH<2 Need 5X	JC/MD	Dilution
21	O2742-01	14-051023	VX035596.D	10 May 2023 19:06	vial A pH<2 Hit of Bromodichloromethane, high in CC	JC/MD	ReRun
22	O2652-04	RO-290535	VX035597.D	10 May 2023 19:29	vial A pH#5.0	JC/MD	Ok
23	VSTDCCC050	VSTDCCC050EC	VX035598.D	10 May 2023 19:52		JC/MD	Ok,M

M : Manual Integration