

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY042423

Review By	Krupa Patel	Review On	4/25/2023 2:57:10 PM				
Supervise By	Mahesh Dadoda	Supervise On	4/25/2023 2:58:19 PM				
SubDirectory	VY042423	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042423s.m		
STD. NAME		STD REF.#					
Tune/Reschk		VP120034,VP120070					
Initial Calibration Stds		VP120035,VP120036,VP120037,VP120038,VP120039,VP120040					
CCC		VP120042,VP120068,VP120069					
Internal Standard/PEM		VP116738					
ICV/I.BLK		VP120042					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY013391.D	24 Apr 2023 09:12		KP/MD	Ok
2	VSTDICV050	ICVVY042423	VY013392.D	24 Apr 2023 10:23		KP/MD	Not Ok
3	VSTDICC005	VSTDICC005	VY013393.D	24 Apr 2023 11:26	Method Created for DOD	KP/MD	Ok,M
4	VSTDICC010	VSTDICC010	VY013394.D	24 Apr 2023 11:48	Com# 11, 13, 20 are on LR	KP/MD	Ok,M
5	VSTDICC020	VSTDICC020	VY013395.D	24 Apr 2023 12:11		KP/MD	Ok,M
6	VSTDICCC050	VSTDICCC050	VY013396.D	24 Apr 2023 12:34		KP/MD	Ok,M
7	VSTDICC100	VSTDICC100	VY013397.D	24 Apr 2023 12:56		KP/MD	Ok,M
8	VSTDICC150	VSTDICC150	VY013398.D	24 Apr 2023 13:19		KP/MD	Ok,M
9	VIBLK	VIBLK	VY013399.D	24 Apr 2023 13:43		KP/MD	Ok
10	VSTDICV050	ICVVY042423	VY013400.D	24 Apr 2023 14:26		KP/MD	Ok,M
11	VY0424SBL01	VY0424SBL01	VY013401.D	24 Apr 2023 15:24		KP/MD	Ok
12	VY0424SBS01	VY0424SBS01	VY013402.D	24 Apr 2023 15:52		KP/MD	Ok,M
13	VY0424SBSD01	VY0424SBSD01	VY013403.D	24 Apr 2023 16:15		KP/MD	Ok,M
14	O2406-01RE	SB-12-6.0-6.5RE	VY013404.D	24 Apr 2023 16:40	Vial B, Not Purge	KP/MD	Not Ok
15	O2406-03	SB-13-6.5-7.0	VY013405.D	24 Apr 2023 17:03	Vial B, Internal standard fail	KP/MD	Ok
16	O2371-01	SB-26-9.0-9.5	VY013406.D	24 Apr 2023 17:26	Vial C, Not Purge	KP/MD	Not Ok
17	O2371-03	SB-23-5.0-5.5	VY013407.D	24 Apr 2023 17:50	Vial C, Not Purge	KP/MD	Not Ok
18	O2371-05RE	SB-22-8.5-9.0RE	VY013408.D	24 Apr 2023 18:13	Vial C, Internal standard fail	KP/MD	Confirms
19	O2371-07	SB-21-8.5-9.0	VY013409.D	24 Apr 2023 18:37	Vial C, Internal standard fail	KP/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY042423

Review By	Krupa Patel	Review On	4/25/2023 2:57:10 PM				
Supervise By	Maresh Dadoda	Supervise On	4/25/2023 2:58:19 PM				
SubDirectory	VY042423	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042423s.m		
STD. NAME	STD REF.#						
Tune/Reschk	VP120034,VP120070						
Initial Calibration Stds	VP120035,VP120036,VP120037,VP120038,VP120039,VP120040						
CCC	VP120042,VP120068,VP120069						
Internal Standard/PEM	VP116738						
ICV/I.BLK	VP120042						
Surrogate Standard							
MS/MSD Standard							
LCS Standard							
20	O2371-09RE	SB-20-8.0-8.5RE	VY013410.D	24 Apr 2023 19:00	Vial C, Internal standard fail	KP/MD	Confirms
21	O2371-11RE	SB-19-8.0-8.5RE	VY013411.D	24 Apr 2023 19:24	Vial C, Internal standard fail	KP/MD	Confirms
22	O2371-15RE	SB-17-8.0-8.5RE	VY013412.D	24 Apr 2023 19:47	Vial C, Internal standard fail	KP/MD	Confirms
23	O2371-17	SB-16-9.0-9.5	VY013413.D	24 Apr 2023 20:10	Vial C internal and surr fail	KP/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VY013414.D	24 Apr 2023 20:33		KP/MD	Ok,M
25	BFB	BFB	VY013415.D	24 Apr 2023 22:49		KP/MD	Ok
26	VSTDCCC050	VSTDCCC050	VY013416.D	24 Apr 2023 23:34		KP/MD	Ok,M
27	VY0424SBL02	VY0424SBL02	VY013417.D	25 Apr 2023 00:21		KP/MD	Ok
28	VY0424SBS02	VY0424SBS02	VY013418.D	25 Apr 2023 00:44		KP/MD	Ok,M
29	VY0424SBSD02	VY0424SBSD02	VY013419.D	25 Apr 2023 01:29	Com#73 recovery fail	KP/MD	Ok,M
30	O2427-01	20230264-18-2	VY013420.D	25 Apr 2023 01:54	Vial A	KP/MD	Ok
31	O2427-02	20230287-24-1	VY013421.D	25 Apr 2023 02:17	Vial A	KP/MD	Ok
32	O2427-03	20230306-28-4	VY013422.D	25 Apr 2023 02:40	Vial A Not Purge	KP/MD	Not Ok
33	O2427-04	20230314-30-4	VY013423.D	25 Apr 2023 03:04	Vial A	KP/MD	Ok
34	O2389-14	SPB-4WC-23-04-17	VY013424.D	25 Apr 2023 03:27	Vial A	KP/MD	Ok
35	O2389-18	SPB-5WC-23-04-17	VY013425.D	25 Apr 2023 03:51	Vial A	KP/MD	Ok
36	O2389-22	SPB-6WC-23-04-17	VY013426.D	25 Apr 2023 04:14	Vial A	KP/MD	Ok
37	O2389-26	SPB-7WC-23-04-17	VY013427.D	25 Apr 2023 04:38	Vial A	KP/MD	Ok
38	O2389-30	SPB-8WC-23-04-17	VY013428.D	25 Apr 2023 05:01	Vial A Internal standard fail	KP/MD	ReRun
39	O2423-02	SPD-1WC-23-04-18	VY013429.D	25 Apr 2023 05:24	Vial A Internal standard fail	KP/MD	ReRun

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY042423

Review By	Krupa Patel	Review On	4/25/2023 2:57:10 PM				
Supervise By	Maresh Dadoda	Supervise On	4/25/2023 2:58:19 PM				
SubDirectory	VY042423	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042423s.m		
STD. NAME	STD REF.#						
Tune/Reschk	VP120034,VP120070						
Initial Calibration Stds	VP120035,VP120036,VP120037,VP120038,VP120039,VP120040						
CCC	VP120042,VP120068,VP120069						
Internal Standard/PEM	VP116738						
ICV/I.BLK	VP120042						
Surrogate Standard							
MS/MSD Standard							
LCS Standard							
40	O2423-06	SPD-2WC-23-04-18	VY013430.D	25 Apr 2023 05:48	Vial A	KP/MD	Ok
41	O2423-10	SPE-1WC-23-04-18	VY013431.D	25 Apr 2023 06:11	Vial A	KP/MD	Ok
42	O2423-14	SPE-2WC-23-04-18	VY013432.D	25 Apr 2023 06:35	Vial A	KP/MD	Ok
43	O2390-02	SPC-1WC-23-04-17	VY013433.D	25 Apr 2023 06:58	Vial A	KP/MD	Ok
44	O2390-06	SPC-2WC-23-04-17	VY013434.D	25 Apr 2023 07:22	Vial A, Internal standard fail	KP/MD	ReRun
45	VSTDCCC050	VSTDCCC050EC	VY013435.D	25 Apr 2023 07:44		KP/MD	Ok,M

M : Manual Integration