

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050323

Review By	Mahesh Dadoda	Review On	5/4/2023 2:01:40 PM				
Supervise By	Krupa Patel	Supervise On	5/4/2023 2:09:49 PM				
SubDirectory	VY050323	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y042423S.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP120322					
		VP120320,VP120321					
		VP116738					
Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY013568.D	03 May 2023 09:07		KP/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY013569.D	03 May 2023 09:38		KP/MD	Ok,M
3	VY0503SBL01	VY0503SBL01	VY013570.D	03 May 2023 10:43		KP/MD	Ok
4	VY0503SBS01	VY0503SBS01	VY013571.D	03 May 2023 11:05		KP/MD	Ok,M
5	VY0503SBSD01	VY0503SBSD01	VY013572.D	03 May 2023 11:28		KP/MD	Ok,M
6	O2568-01	72-11938	VY013573.D	03 May 2023 11:52	Vial B, Not purged well	KP/MD	Not Ok
7	O2569-01	ETGI-339	VY013574.D	03 May 2023 12:16	Vial B	KP/MD	Ok
8	O2558-07	P001-SS013-0612-01	VY013575.D	03 May 2023 12:39	Vial A	KP/MD	Ok
9	O2558-08	P001-SS013-1218-01	VY013576.D	03 May 2023 13:03	Vial A	KP/MD	Ok
10	O2558-06	P001-SS013-0006-01	VY013577.D	03 May 2023 13:26	Vial A	KP/MD	Ok
11	O2554-01RE	TR-04-42823RE	VY013578.D	03 May 2023 13:49	Vial A ISTD Fail	KP/MD	Confirms
12	VIBLK	VIBLK	VY013579.D	03 May 2023 14:13		KP/MD	Ok
13	O2500-24	PT-VOA-SOIL	VY013580.D	03 May 2023 14:36	Need 4X(NJ-Soil-PT)	KP/MD	Dilution
14	VIBLK	VIBLK	VY013581.D	03 May 2023 15:00		KP/MD	Ok
15	O2567-02RE	B-21-SB02RE	VY013582.D	03 May 2023 15:23	Vial B Internal standard Fail	KP/MD	Confirms
16	O2500-24DL	PT-VOA-SOILDL	VY013583.D	03 May 2023 15:46		KP/MD	Ok
17	O2573-02	VNJ-240	VY013584.D	03 May 2023 16:11	Vial A, Surrogate Fail, Internal standard Fail	KP/MD	ReRun
18	O2582-01	HD-02-050223	VY013585.D	03 May 2023 16:34	Vial A	KP/MD	Ok
19	O2584-01	TR-05-050123	VY013586.D	03 May 2023 16:57	Vial A	KP/MD	Ok



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050323

Review By	Mahesh Dadoda	Review On	5/4/2023 2:01:40 PM				
Supervise By	Krupa Patel	Supervise On	5/4/2023 2:09:49 PM				
SubDirectory	VY050323	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y042423S.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		VP120322					
CCC		VP120320,VP120321					
Internal Standard/PEM		VP116738					
ICV/I.BLK							
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

20	O2585-01	OK-01-050223	VY013587.D	03 May 2023 17:21	Vial A	KP/MD	Ok
21	O2571-01	OE-3-050123	VY013588.D	03 May 2023 17:44	Vial A	KP/MD	Ok
22	O2559-02	VOC	VY013589.D	03 May 2023 18:08	Vial A	KP/MD	Ok
23	O2505-04	WO-4	VY013590.D	03 May 2023 18:31	Vial A, Internal standard Fail	KP/MD	ReRun
24	VSTDCCC050	VSTDCCC050EC	VY013591.D	03 May 2023 18:54		KP/MD	Ok,M

M : Manual Integration