

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050423

Review By	Krupa Patel	Review On	5/5/2023 11:53:24 AM				
Supervise By	Mahesh Dadoda	Supervise On	5/5/2023 11:54:06 AM				
SubDirectory	VY050423	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y042423S.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP120322					
		VP120323,VP120324					
		VP116738					
Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY013592.D	04 May 2023 09:17		KP/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY013593.D	04 May 2023 10:51		KP/MD	Ok,M
3	VY0504SBL01	VY0504SBL01	VY013594.D	04 May 2023 12:56		KP/MD	Ok
4	VY0504SBS01	VY0504SBS01	VY013595.D	04 May 2023 13:18		KP/MD	Ok,M
5	VY0504SBSD01	VY0504SBSD01	VY013596.D	04 May 2023 13:41		KP/MD	Ok,M
6	O2573-02RE	VNJ-240RE	VY013597.D	04 May 2023 14:05	Vial B, Internal standard Fail	KP/MD	Confirms
7	O2505-05	WO-5	VY013598.D	04 May 2023 14:29	Vial A, Internal standard Fail, Need MeOH	KP/MD	Dilution
8	O2598-04	ETGI-328	VY013599.D	04 May 2023 14:52	Vial A, Surrogate Fail, ISTD Fail	KP/MD	Ok
9	VIBLK	VIBLK	VY013600.D	04 May 2023 15:16		KP/MD	Ok
10	O2598-01	VNJ-241	VY013601.D	04 May 2023 15:39	Vial A, Internal standard Fail	KP/MD	Ok
11	O2607-01	HT-1	VY013602.D	04 May 2023 16:02	Vial A	KP/MD	Ok
12	O2605-01	RC-1	VY013603.D	04 May 2023 16:26	Vial A Not Purge	KP/MD	Not Ok
13	O2604-01	TF-1	VY013604.D	04 May 2023 16:49	Vial A Internal standard Fail	KP/MD	Ok
14	O2505-04RE	WO-4RE	VY013605.D	04 May 2023 17:13	Vial B Internal standard Fail	KP/MD	Confirms
15	O2588-05	ZONE-C-3-6-05012023	VY013606.D	04 May 2023 17:36	Vial A	KP/MD	Ok
16	O2588-08	ZONE-A-3-6-05022023	VY013607.D	04 May 2023 18:00	Vial A	KP/MD	Ok
17	O2588-10	ZONE-D-0-3-05022023	VY013608.D	04 May 2023 18:23	Vial A	KP/MD	Ok
18	O2588-11	ZONE-D-3-6-05022023	VY013609.D	04 May 2023 18:46	Vial A	KP/MD	Ok
19	O2588-15	PILE-B-UNK-05022023	VY013610.D	04 May 2023 19:10	Vial A	KP/MD	Ok



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050423

Review By	Krupa Patel	Review On	5/5/2023 11:53:24 AM					
Supervise By	Maresh Dadoda	Supervise On	5/5/2023 11:54:06 AM					
SubDirectory	VY050423	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y042423S.M			
STD. NAME		STD REF.#						
Tune/Reschk Initial Calibration Stds		VP120322						
CCC		VP120323,VP120324						
Internal Standard/PEM		VP116738						
ICV/I.BLK								
Surrogate Standard								
MS/MSD Standard								
LCS Standard								
20	O2588-06	ZONE-C-6-9-05012023	VY013611.D	04 May 2023 19:33	Vial A	KP/MD	Ok	
21	VSTDCCC050	VSTDCCC050EC	VY013612.D	04 May 2023 19:56		KP/MD	Ok,M	

M : Manual Integration