



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: FID\_E

**Daily Analysis Runlog For Sequence/QC Batch ID # FE042123AL**

|                                         |                                         |                   |                                 |
|-----------------------------------------|-----------------------------------------|-------------------|---------------------------------|
| Review By                               | yogesh                                  | Review On         | 4/20/2023 3:09:59 PM            |
| Supervise By                            | Ankita                                  | Supervise On      | 4/21/2023 8:46:38 AM            |
| SubDirectory                            | FE042123AL                              | HP Acquire Method | HP Processing Method FE042123AL |
| <b>STD. NAME</b>                        | <b>STD REF.#</b>                        |                   |                                 |
| Tune/Reschk<br>Initial Calibration Stds | PP21954,PP21956,PP21957,PP21958,PP21959 |                   |                                 |
| CCC                                     | PP21957                                 |                   |                                 |
| Internal Standard/PEM                   |                                         |                   |                                 |
| ICV/I.BLK                               | PP21955,PP21960                         |                   |                                 |
| Surrogate Standard                      |                                         |                   |                                 |
| MS/MSD Standard                         |                                         |                   |                                 |
| LCS Standard                            |                                         |                   |                                 |

| Sr# | SampleID                    | Data File Name | Date-Time         | Operator | Status |
|-----|-----------------------------|----------------|-------------------|----------|--------|
| 1   | MECL2                       | FE041454.D     | 20 Apr 2023 16:53 | YP\AJ    | Ok     |
| 2   | I.BLK                       | FE041455.D     | 20 Apr 2023 17:23 | YP\AJ    | Ok     |
| 3   | 100 PPM ALIPHATIC HC STD1   | FE041456.D     | 20 Apr 2023 17:54 | YP\AJ    | Ok     |
| 4   | 50 PPM ALIPHATIC HC STD2    | FE041457.D     | 20 Apr 2023 18:24 | YP\AJ    | Ok     |
| 5   | 20 PPM ALIPHATIC HC STD3    | FE041458.D     | 20 Apr 2023 18:54 | YP\AJ    | Ok     |
| 6   | 10 PPM ALIPHATIC HC STD4    | FE041459.D     | 20 Apr 2023 19:24 | YP\AJ    | Ok,M   |
| 7   | 5 PPM ALIPHATIC HC STD5     | FE041460.D     | 20 Apr 2023 19:55 | YP\AJ    | Ok,M   |
| 8   | 20 PPM ALIPHATIC HC STD ICV | FE041461.D     | 20 Apr 2023 20:25 | YP\AJ    | Ok     |
| 9   | I.BLK                       | FE041462.D     | 20 Apr 2023 21:55 | YP\AJ    | Ok     |
| 10  | 20 PPM ALIPHATIC HC STD     | FE041463.D     | 20 Apr 2023 22:26 | YP\AJ    | Ok,M   |

M : Manual Integration