



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE042723AL

Review By	yogesh	Review On	4/27/2023 11:01:52 AM				
Supervise By	Ankita	Supervise On	5/1/2023 9:06:27 AM				
SubDirectory	FE042723AL	HP Acquire Method	HP Processing Method		FE042123AL		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		PP21954,PP21956,PP21957,PP21958,PP21959					
CCC		PP21957					
Internal Standard/PEM		PP21955,PP21960					
ICV/I.BLK							
Surrogate Standard							
MS/MSD Standard							
LCS Standard							
Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FE041562.D	27 Apr 2023 07:50		YPIAJ	Ok
2	I.BLK	I.BLK	FE041563.D	27 Apr 2023 08:20		YPIAJ	Ok
3	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE041564.D	27 Apr 2023 08:50		YPIAJ	Ok
4	O2491-03	S2S-TP4	FE041565.D	27 Apr 2023 09:21		YPIAJ	Ok
5	I.BLK	I.BLK	FE041566.D	27 Apr 2023 09:51		YPIAJ	Ok
6	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE041567.D	27 Apr 2023 10:51		YPIAJ	Ok
7	O2503-07	TR-305A-03	FE041568.D	27 Apr 2023 11:39		YPIAJ	Ok
8	O2503-12	TR-305A-04	FE041569.D	27 Apr 2023 12:10		YPIAJ	Ok
9	O2503-19	TR-305A-07	FE041570.D	27 Apr 2023 12:40		YPIAJ	Ok
10	O2503-20	TR-305A-08	FE041571.D	27 Apr 2023 13:10		YPIAJ	Ok
11	O2503-21	TR-305A-09	FE041572.D	27 Apr 2023 13:40		YPIAJ	Ok
12	O2504-05	BU-512-01	FE041573.D	27 Apr 2023 14:11	surrogate fail	YPIAJ	Not Ok
13	O2504-06	BU-512-02	FE041574.D	27 Apr 2023 14:41		YPIAJ	Ok
14	O2504-07	BU-512-03	FE041575.D	27 Apr 2023 15:11	surrogate fail	YPIAJ	Not Ok
15	O2504-12	BU-512-04	FE041576.D	27 Apr 2023 15:41		YPIAJ	Ok
16	O2504-13	BU-512-05	FE041577.D	27 Apr 2023 16:11		YPIAJ	Ok
17	O2504-14	BU-512-06	FE041578.D	27 Apr 2023 16:41		YPIAJ	Ok
18	O2504-19	BU-512-07	FE041579.D	27 Apr 2023 17:12		YPIAJ	Ok
19	O2504-20	BU-512-08	FE041580.D	27 Apr 2023 17:42		YPIAJ	Ok

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE042723AL

Review By	yogesh	Review On	4/27/2023 11:01:52 AM					
Supervise By	Ankita	Supervise On	5/1/2023 9:06:27 AM					
SubDirectory	FE042723AL	HP Acquire Method	HP Processing Method		FE042123AL			
STD. NAME		STD REF.#						
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		PP21954,PP21956,PP21957,PP21958,PP21959						
		PP21957						
		PP21955,PP21960						
20	O2504-21	BU-512-09	FE041581.D	27 Apr 2023 18:12		YPIAJ	Ok	
21	O2507-01	HD-01-042623	FE041582.D	27 Apr 2023 18:42		YPIAJ	Ok	
22	O2507-02	HD-01-042623-EPH-2	FE041583.D	27 Apr 2023 19:12		YPIAJ	Ok	
23	O2503-13	TR-305A-05	FE041584.D	27 Apr 2023 19:42	need 10x dilution	YPIAJ	Dilution	
24	O2503-14	TR-305A-06	FE041585.D	27 Apr 2023 20:13	need 20x dilution	YPIAJ	Dilution	
25	O2503-13DL	TR-305A-05DL	FE041586.D	27 Apr 2023 20:43		YPIAJ	Ok	
26	O2503-14DL	TR-305A-06DL	FE041587.D	27 Apr 2023 21:13	Need 20X	YPIAJ	Not Ok	
27	I.BLK	I.BLK	FE041588.D	27 Apr 2023 22:13		YPIAJ	Ok	
28	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE041589.D	27 Apr 2023 23:44		YPIAJ	Ok,M	
29	O2502-05	BU-511-01	FE041590.D	28 Apr 2023 00:15		YPIAJ	Ok	
30	O2502-06	BU-511-02	FE041591.D	28 Apr 2023 00:44		YPIAJ	Ok	
31	O2502-07	BU-511-03	FE041592.D	28 Apr 2023 01:14		YPIAJ	Ok	
32	O2502-12	BU-511-04	FE041593.D	28 Apr 2023 01:45		YPIAJ	Ok	
33	O2505-01	WO-1	FE041594.D	28 Apr 2023 02:15		YPIAJ	Ok	
34	O2505-02	WO-2	FE041595.D	28 Apr 2023 02:45		YPIAJ	Ok	
35	O2509-01	HL-1	FE041596.D	28 Apr 2023 03:15		YPIAJ	Ok	
36	O2509-02	O2509-02	FE041597.D	28 Apr 2023 03:46	surrogate fail	YPIAJ	ReRun	
37	O2509-03	O2509-03	FE041598.D	28 Apr 2023 04:16	surrogate fail	YPIAJ	ReRun	
38	O2509-04	HL-4	FE041599.D	28 Apr 2023 04:46		YPIAJ	Ok	
39	O2509-05	HL-5	FE041600.D	28 Apr 2023 05:16		YPIAJ	Ok	

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE042723AL

Review By	yogesh	Review On	4/27/2023 11:01:52 AM				
Supervise By	Ankita	Supervise On	5/1/2023 9:06:27 AM				
SubDirectory	FE042723AL	HP Acquire Method	HP Processing Method		FE042123AL		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		PP21954,PP21956,PP21957,PP21958,PP21959					
CCC		PP21957					
Internal Standard/PEM		PP21955,PP21960					
ICV/I.BLK							
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

40	O2509-06	HL-6	FE041601.D	28 Apr 2023 05:46		YPIAJ	Ok
41	O2509-07	HL-7	FE041602.D	28 Apr 2023 06:17		YPIAJ	Ok
42	O2505-03	WO-3	FE041603.D	28 Apr 2023 06:47	need 20x dilution	YPIAJ	Dilution
43	O2505-04	WO-4	FE041604.D	28 Apr 2023 07:17	need 25x dilution	YPIAJ	Dilution
44	O2505-06	WO-6	FE041605.D	28 Apr 2023 07:47	need 20x dilution	YPIAJ	Dilution
45	O2505-05	WO-5	FE041606.D	28 Apr 2023 08:17	need 100x dilution	YPIAJ	Dilution
46	I.BLK	I.BLK	FE041607.D	28 Apr 2023 09:17		YPIAJ	Ok
47	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE041608.D	28 Apr 2023 09:48		YPIAJ	Ok,M
48	O2504-05	BU-512-01	FE041609.D	28 Apr 2023 11:19	Use this run	YPIAJ	Ok
49	O2504-07	BU-512-03	FE041610.D	28 Apr 2023 11:49	Use this run	YPIAJ	Ok
50	O2507-01DL	HD-01-042623DL	FE041611.D	28 Apr 2023 12:19	Not required	YPIAJ	Not Ok
51	O2507-02DL	HD-01-042623-EPH-2D	FE041612.D	28 Apr 2023 12:50	Not required	YPIAJ	Not Ok
52	O2505-03DL	O2505-03DL	FE041613.D	28 Apr 2023 13:20	Need higher dilution	YPIAJ	Not Ok
53	O2505-04DL	O2505-04DL	FE041614.D	28 Apr 2023 13:50	Need higher dilution	YPIAJ	Not Ok
54	O2505-06DL	O2505-06DL	FE041615.D	28 Apr 2023 14:21	Need higher dilution	YPIAJ	Not Ok
55	O2505-05DL	O2505-05DL	FE041616.D	28 Apr 2023 14:50	Need higher dilution	YPIAJ	Not Ok
56	O2509-02RE	O2509-02RE	FE041617.D	28 Apr 2023 15:21	possible carry over, surrogate fail	YPIAJ	Not Ok
57	O2509-03RE	O2509-03RE	FE041618.D	28 Apr 2023 15:51	possible carry over, surrogate fail	YPIAJ	Not Ok
58	I.BLK	I.BLK	FE041619.D	28 Apr 2023 16:52		YPIAJ	Ok
59	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE041620.D	28 Apr 2023 17:22		YPIAJ	Ok,M



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE042723AL

Review By	yogesh	Review On	4/27/2023 11:01:52 AM
Supervise By	Ankita	Supervise On	5/1/2023 9:06:27 AM
SubDirectory	FE042723AL	HP Acquire Method	HP Processing Method FE042123AL
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP21954,PP21956,PP21957,PP21958,PP21959		
CCC Internal Standard/PEM	PP21957		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP21955,PP21960		

M : Manual Integration