



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC042723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063465.D

Aliphatic C9-C12	8072902.000	60.000	3.199	6.486	134548.367	137065.155	1.836
Aliphatic C12-C16	5303673.000	40.000	6.487	9.878	132591.825	135585.936	2.208
Aliphatic C16-C21	8091788.000	60.000	9.879	13.234	134863.133	135817.180	0.702
Aliphatic C21-C28	10622590.000	80.000	13.235	16.884	132782.375	134266.069	1.105
Aliphatic C28-C40	15339507.000	120.000	16.885	21.741	127829.225	131023.808	2.438
Aliphatic EPH	47430460.000	360.000	3.199	21.741	131751.278	134057.000	1.720