



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC042723AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063475.D

Aliphatic C9-C12	8136702.000	60.000	3.199	6.486	135611.700	137065.155	1.060
Aliphatic C12-C16	5374347.000	40.000	6.487	9.878	134358.675	135585.936	0.905
Aliphatic C16-C21	8207459.000	60.000	9.879	13.234	136790.983	135817.180	-0.717
Aliphatic C21-C28	11048644.000	80.000	13.235	16.884	138108.050	134266.069	-2.861
Aliphatic C28-C40	17114924.000	120.000	16.885	21.741	142624.367	131023.808	-8.854
Aliphatic EPH	49882076.000	360.000	3.199	21.741	138561.322	134057.000	-3.360