



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC050123AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC063488.D

Aliphatic C9-C12	9589554.000	60.000	3.198	6.484	159825.900	137065.155	-16.606
Aliphatic C12-C16	6366856.000	40.000	6.485	9.880	159171.400	135585.936	-17.395
Aliphatic C16-C21	9694428.000	60.000	9.881	13.239	161573.800	135817.180	-18.964
Aliphatic C21-C28	12856959.000	80.000	13.240	16.891	160711.988	134266.069	-19.697
Aliphatic C28-C40	16762510.000	120.000	16.892	21.751	139687.583	131023.808	-6.612
Aliphatic EPH	55270307.000	360.000	3.198	21.751	153528.631	134057.000	-14.525